CORNING

Certificate of Compliance

836 North Street - Life Sciences
Tewksbury MA 01876 USA www.corning.com/lifesciences
Refer to website for regional contact information.

ul. Geodetów 23
84-120 Władysławowo
www.eproscience.com

Product Name:	***Corning Axygen® Universal Filter Pipette Tips***
Catalog Number:	***PRFL-300-C, PRFL-200-C***
Lot ID:	***.....***
Expiration Date :	***.....***
Manufacture Date:	***.....***

Quality Management System - Complies with the current version of the ISO 9001 Standard***.***

Non-Pyrogenic - Tested and met the criteria established in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing" and USP <85>, "Bacterial Endotoxins Test". The acceptance level for product is ≤ 0.06 EU/ml.

BSE/TSE - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via humanand veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

Human DNA Free - Tested by PCR method and found to be free of detectable human DNA contamination. The assay detection limit is hDNA 5pg.

DNase/RNase Free - Tested by nuclease assay method and found to be free of detectable DNase/RNase contamination. The assay detection limit is 10^{-7} Kunitz units/ul for DNase and 10^{-9} Kunitz units/ul for RNase.

ATP Free - Tested to be <10^{-13} mg/ul for detectible levels of Adenosine-5'-triphospate (ATP).

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass
Dimensional Inspection - Pass
Functional Test - Pass
Accuracy Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

For Certificate of Compliance for specific LOT number,
please contact us at the following email address: tenders@eproscience.com.